

# ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0508D

System: Skyrad PSPG

Site: Darwin

## OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 <sup>-6</sup> or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
PSP	31285	Skyrad	9/2/2004	8.1777	na	8/5/2005	00:00	1:30

## NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 <sup>-6</sup> or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
PSP	31291	Skyrad	8/5/2005	8.389	na	X

Replacement completed by: Gary Eckert, Mike Alsop, Bill  
Porch

## NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP

Other old part disposition (explain):

Other comments: